

Manual Supplement

Manual Title: 707Ex Users
Print Date: July 2004
Revision/Date: 4, 8/09

Supplement Issue: **3**
Issue Date: 9/17
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 707Ex
CD Rev. & Date: 4, 8/2009
CD PN: 2053979

Change #1, 52974

On page 13, under **Ex-Certification Data**, replace the second bullet with the following:

- Certification:  II 2 G Ex ia IIC T4

Change #2, 58414

On page 14, add the following under **source / Simulate mA dc**:

Where in a 3V/m radiated EM field of 300 MHz-500 MHz, floor counts are increased to 40 counts in mA DC source/simulate mode.

Change #3, 514

On page 13, under **Approved Batteries** and before the table, add:

Important Intrinsic Safety component update

The Panasonic Powerline Industrial 9V battery model 6LR61.9V has been reevaluated and found no longer suitable for use in intrinsically safe products. At this time only the batteries as detailed below have been found to be compliant. Please replace the Approved battery table as follows:

Also on page 13, replace the **Approved Batteries** table with:

Battery	Manufacturer	Type
Alkaline, 9 volt	Duracell	6LR61/MN1604
Alkaline Energizer, 9 volt	Eveready	6LR61/522