

# Supplement

Title:	700LTP-1 I.S.	Supplement Issue:	1
Part Number:	2811813	Issue Date:	3/11
Print Date:	April 2007	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

## Change #1, 57805

Under ***Introduction***,

Change: Output vacuum: 0 to -13 psi / 0 to -900 mbar

To: Output vacuum: 0 to -12.7 psi / 0 to -875 mbar